

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate No.
R60/2000-DE1-10.10

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Dirk Ratschko

Applicant

Name: Keli Electric Manufacturing (Ningbo) Co. Ltd.
Address: NO. 199 Changxing Road
315033 Ningbo, Jiangbei District
China

Manufacturer of the certified type is the applicant.

Identification of the certified type Strain gauge double bending beam load cell

Type: ILY-SS

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60, edition 2000
for accuracy class C3

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.


OIML Certificate No.
R60/2000-DE1-10.10

The conformity was established by the results of tests and examinations provided in the associated Test Report and Report

Test Report No. 4048383-1
Report No. 4048383

that includes 22 pages
that includes 6 pages

The Issuing Authority




Dr. D. Ratschko
Head of Department

06.10.2010



The OIML Member



Dr. R. Schwartz
Head of Division

06.10.2010

The load cells of the series ILY-SS are double bending beam load cells. They are made of stainless steel and the strain gauge application is hermetically sealed. The metrological characteristics for application in approved weighing instruments are listed in table 1.

Table 1: Essential data

Accuracy class			C3
Maximum number of load cell intervals	n_{LC}		3000
Rated output		mV/V	2
Maximum capacity	E_{max}	kg	8 / 10 / 15 / 20 / 30 / 40
Minimum load cell verification interval	$\frac{V_{min}}{(E_{max} / Y)}$		$E_{max} / 8500$

Dead load: $0\% \cdot E_{max}$; Safe overload: $150\% \cdot E_{max}$; Input impedance: 400Ω ; Fraction: $p_{LC} = 0.7$

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.