

General Administration of Quality Supervision, Inspection  
and Quarantine of the People's Republic of China

OIML Certificate No.  
R60/2000-CN-03.03



# OIML CERTIFICATE OF CONFORMITY

To Keli Sensor Manufacturing(Ningbo) Co., LTD.

This certificate attests the conformity of the type of the Load

Cell SB with the requirements

of the Recommendation R60(2000) of the International

Organization of Legal Metrology(OIML).

The issuing authority  
*Office of OIML Affairs*



*May 14, 2003*

The OIML member

王春平

*May 14, 2003*

Member State  
P.R.China

OIML Certificate No.  
R60/2000-CN-03.03

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: No.9 Madiandonglu, Haidian District, Beijing, China

## Applicant

Name: Keli Sensor Manufacturing(Ningbo)Co. Ltd.  
Address: No.181, Canghai Road, Hi-Tech Industrial Park, Ningbo City 315040 P.R. China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:  
Address:

## Identification of the certified type

Model designation	SB			
Accuracy class	°C3			
Maximum capacity ( $E_{max}$ )	3t	5t	10t	15t
Minimum dead load( $E_{min}$ )	0	200kg	400kg	600kg
Maximum number of load cell intervals ( $n_{max}$ )	3000			
Minimum load cell verification interval ( $V_{min}$ )	200g	1000g	2000g	3000g
Apportionment factor( $P_{LC}$ )	0.7			
Safe overload	3.6t	6t	12t	18t
Operating temperature range	-10°C - +40°C			
Humidity condition	CH			

The numbers of the associated test reports and the included pages:

LXff2003-6002 18 pages

OIML Certificate No.  
R60/2000-CN-03.03

**This certificate attests the conformity of the above-mentioned type (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).**

R: 60

Edition: 2000

For accuracy class (if applicable): C3

**This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned, as covered by the relevant OIML International Recommendation(s).**

**This certificate does not bestow any form of legal international approval.**

**The conformity was established by tests described in the associated test reports.**

The issuing authority

Office of OIML Affairs

中国秘书处

May 14, 2003

The CIML member

王秦平

May 14, 2003

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.